


10X Brilliant Assay Buffer

Lot 10812a

Method	Specification	Analysis
Description		
Appearance	Clear, Colorless Solution	Conforms
pH	Mettler Toledo SevenCompact S-210 pH Meter	
pH of 1X Solution ¹	7.30 ± 0.10	7.28
Osmolality	Wescor 5520 Vapor Pressure Osmometer	
Osmolality of 1X Solution ¹	305 ± 15 mmol/kg	300
Cell Assay	Molecular Devices FlexStation 3 Multi-mode Plate Reader	
Used as part of a dye-loading solution in a Cell Assay ²	The assay returns a passing result for the selected Dye Indicator used in the corresponding bio-validation method	Passes

¹Diluted to 1X concentration with Type II Lab-Grade Water; ²Bio-Validation Method: Fluo-4 AM Cell Validation



David Woody

Quality Manager

3 Feb 2026