


10X Brilliant Chloride Free Stimulus Buffer

Lot 10612a

Method	Specification	Analysis
Description		
Appearance	Clear, Colorless to Yellow Solution	Clear, Yellow
pH	Mettler Toledo SevenCompact S-210 pH Meter	
pH of 1X Solution ¹	7.25 ± 0.1	7.22
Osmolality	Wescor 5500 Vapor Pressure Osmometer	
Osmolality of 1X Solution ¹	290 ± 15 mmol/kg	289
Cell Assay	BioTek Cytation 5 Imaging Plate Reader	
Used as part of a Stimulus solution in a Cell Assay ²	The assay returns a passing result for the selected Dye Indicator used in the corresponding bio-validation method	Passes

¹Diluted to 1X concentration with Type II Lab-Grade Water; ²Bio-validation Method: Thallos AM Cell Validation



David Woody

Quality Manager

1 Nov 2024