




BAPTA JF549™ AM

Lot 10621a

Method	Specification	Analysis
LCMS	Agilent 1220 Infinity II	
Purity ¹	≥ 90%	93.7%
Molecular Ion ²	<i>Common Peaks</i> 1216.4 ± 0.5 m/z [M+H] ⁺ 608.7 ± 0.5 m/z [M+2H] ²⁺	<i>Detected Peaks</i> 1216.8 m/z 609.1 m/z
Absorbance Spectrum	Agilent Cary 60 UV-VIS Spectrophotometer	
UV-Visible λ_{\max} ³	546 ± 3 nm	545 nm
Fluorescence Spectrum	Horiba Jobin Yvon FluoroMax Plus Spectrofluorometer	
Excitation λ_{\max} ³ Emission λ_{\max}	546 ± 3 nm 569 ± 3 nm	545 nm 570 nm
¹H NMR Spectrum	Bruker Avance 400	
Peaks and Integrations	Conforms to Structure	Conforms

¹Column: Phenomenex 00d-4251-E0 Luna C₁₈, 4.6 x 100 mm, 100Å, 3 µm, UV-Vis Diode Array Detector: 254 nm; ²Single Quad MS Detector: ESI Positive; ³Solvent: 10 mM CaEGTA, 100 mM KCl, 10mM MOPS, pH 7.2, AM esters hydrolyzed to ion-sensing salt form prior to acquiring spectral data



David Woody

Quality Manager

20 Dec 2024