




IPG-2 AM

Lot 10613a

Method	Specification	Analysis
LCMS	Agilent 1220 Infinity II	
Purity ¹	≥ 90%	94.0%
Molecular Ion ²	<i>Common Peaks</i> 1127.36 ± 0.5 m/z [MH] ⁺ 564.18 ± 0.5 m/z [MH ₂] ²⁺	<i>Detected Peaks</i> 1127.7 m/z 564.4 m/z
Absorbance Spectrum	Agilent Cary 60 UV-VIS Spectrophotometer	
UV-Visible λ _{max} ³	517 ± 3 nm	518 nm
Fluorescence Spectrum	Horiba Jobin Yvon FluoroMax 4 Spectrofluorometer	
Excitation λ _{max} ³ Emission λ _{max}	517 ± 3 nm 540 ± 3 nm	516 nm 538 nm
¹H NMR Spectrum	Bruker Avance 400	
Peaks and Integrations	Conforms to Structure	Conforms
Cell Assay	BioTek Cytation 5 Imaging Reader	
F/F _{0, Ctrl} - F/F _{0, Stim} post-stimulus in relevant biological assay	≥ 0.26	0.34

¹Column: Phenomenex 00d-4251-E0 Luna C₁₈, 4.6 x 100 mm, 3 μm, UV-Vis Diode Array Detector: 254 nm; ²Single Quad MS Detector: ESI Positive; ³Solvent: 140mM KCl in 10 mM MOPS pH 7.2, AM esters hydrolyzed to ion-sensing salt form prior to acquiring spectral data



David Woody

Quality Manager

20 Aug 2024