



Thallos AM

Lot 10212a

Method	Specification	Analysis
LCMS	Agilent 1220 Infinity II	
Purity*	≥ 95%	95.9 %
Molecular Ion	841.21 ± 0.5 m/z (MH ⁺)	841.5 m/z
Absorbance Spectrum	Agilent Cary 60 UV-VIS Spectrophotometer	
Longest-Wavelength Absorbance Maximum**	494 ± 3 nm	494 nm
Fluorescence Spectrum	Horiba Jobin Yvon FluoroMax 4 Spectrofluorometer	
Excitation Max.; Emission Max.**	490 ± 3 nm; 515 ± 3 nm	491 nm; 516 nm
¹H NMR Spectrum	Bruker Avance 400	
Peaks and Integrations	Only relevant product peaks — with appropriate chemical shifts and peak integrations — and solvent peaks present	Confirmed
Cell Assay	BioTek Cytation 5 Imaging Reader	
F/F ₀ post-stimulus in relevant biological assay	≥ 1.8	2.1

*Column: Agilent Infinity Lab Poroshell 120 ECC18, 30 x 50 mm, 2.7 μm C₁₈, UV-Vis Diode Array Detector: 254 nm, Single Quad MS Detector: ESI Positive; **solvent: MOPS

Approved by P. Rogelio Escamilla May 2023